

Test Sequence

Defibrillator Annual Test (v1.0)

Template	Defibrillator Annual Test
Version	1.0
Created	2025-10-02 00:53
Created By	Seeder

Test Steps

#	Step / Instruction	Type	Limits	Unit	Actual / Pass
1	Delivered Energy @ 200 J Charge to 200 J and discharge into analyzer Ref: IEC 60601-2-4 §201.12.4.101	Measurement	200 J ±15% ±1	J	_____
2	Sync Function Verify R-wave synchronisation within 60 ms Ref: IEC 60601-2-4	Yes/No Check	-	-	_____