

Test Sequence

SpO₂ Monitoring – ISO 80601-2-61 (v1.0)

Template	SpO ₂ Monitoring – ISO 80601-2-61
Version	1.0
Created	2025-10-02 01:08
Created By	Seed

Test Steps

#	Step / Instruction	Type	Limits	Unit	Actual / Pass
1	SpO₂ Accuracy at 98% Simulate 98% with compatible sensor; record value. Ref: ISO 80601-2-61 §201.12.1	Measurement	98 % ±2	%	_____
2	SpO₂ Accuracy at 90% Simulate 90%; record displayed value. Ref: ISO 80601-2-61 §201.12.1	Measurement	90 % ±3	%	_____