

Test Sequence

Medical Bed – IEC 60601-2-52 (v1.0)

Template	Medical Bed – IEC 60601-2-52
Version	1.0
Created	2025-10-02 01:08
Created By	Seed

Test Steps

#	Step / Instruction	Type	Limits	Unit	Actual / Pass
1	Side Rail Integrity & Spacing Inspect rails for secure locking and entrapment spacing. Ref: IEC 60601-2-52 §201.9 & Annex AA	Yes/No Check	-	-	_____
2	Brake & Castor Function Verify brakes hold bed on flat surface; steering castor operates. Ref: IEC 60601-2-52 §201.9.8	Yes/No Check	-	-	_____