

Test Sequence

ECG Monitoring – IEC 60601-2-27 (v1.0)

Template	ECG Monitoring – IEC 60601-2-27
Version	1.0
Created	2025-10-02 01:07
Created By	Seed

Test Steps

#	Step / Instruction	Type	Limits	Unit	Actual / Pass
1	Heart Rate Accuracy (60 bpm) Simulate sinus 60 bpm; record displayed value. Ref: IEC 60601-2-27 §201.12.4.101	Measurement	60 bpm ±1	bpm	_____
2	Arrhythmia Detection Simulate VF/asystole; verify alarms. Ref: IEC 60601-2-27	Yes/No Check	-	-	_____